

ARCS Instrument Replacement Form

Fill out the following form when replacing instruments at an ARCS site:

System: SMET

Site: Nauru

OLD Configuration

Part Number	Serial Number	Location	Configuration Date	Cal Gain Coefficient (Responsivity at 45°) x 10 ⁻⁶ or 4 PIR coefs.	Offset	Date of Change (UTC)	Time Start UTC	Time End UTC
PTB201A	505303	Barometer	V060901.00	na	na	23/Jan/07	00:20	00:30

NEW Configuration

Part Number	Serial Number	Location	Configuration Date	Cal Gain Coefficient (Responsivity at 45°) x 10 ⁻⁶ or 4 PIR coefs.	Offset	Change Logger Coefficients and Version
PTB220	Y1330002	Barometer	V060901.00	na	na	na

Replacement completed by: Troy Culkan

NOTES:

Old Part remains on site as spare:

Old part packed for shipment to:

Other old part disposition (explain): PTB 201 are no longer supported for repair

Other comments: